OPERATIONS MANUAL

with BluWave™ SP



IN-THE-CANAL

Table of Contents

Overview
Features, Controls and Identification2
Preparation
Batteries
Battery Indicator Tones
Insertion and Removal
Operation
On, Off and Volume Levels with No Volume Control 1
On, Off and Volume Levels with Volume Control 12
Multimemory Settings
Directional Microphones
Directional Settings16
Telephone Use18
Self Check
Reminder23
Summary of Settings
Hearing System Care
Instrument Care
Service and Repair
Troubleshooting Guide
Tips for Better Communication
For You30
For Your Family and Friends3
FDA Information
Required Information

Overview


Features, Controls and Identification


Your hearing system controls include:

- 1. Microphone opening(s)
- 2. Battery compartment (on/off control)
- 3. Vent (optional)
- 4. Receiver opening

Your hearing system can be identified by:

5. Location of serial number: RED Right BLUE Left









Batteries


Your hearing system uses either a size 10 or 312 battery as its power source. This battery size can be identified by the yellow (10) or brown (312) color code on the packaging.





To insert or replace the battery:

- 1. Use the nail grip on the battery door.
- 2. Open the battery door gently and remove the old battery.
- 3. Remove the tab from the new battery.
- 4. Line up the battery's plus "+" sign (flat side of the battery) with the "+" on the battery door.
- 5. Close the battery door.



Battery Indicator Tones

An indicator will sound when the battery voltage is low. You have approximately 20 minutes* to replace the battery. A different tone lasting two seconds will sound just before the battery stops working.

* Actual time between low battery indicator and shut down will vary depending on environmental noise levels and brand of battery used.

Helpful Hints

- NEVER FORCE THE DOOR SHUT; this could result in serious damage; if the door will not close securely, check that the battery is inserted correctly
- Do not open the battery door too far or damage is likely to occur
- Because of their size, it's a good idea to change and replace batteries above a table or desk to reduce the risk of dropping the instrument or battery

- Dispose of used batteries immediately in the proper waste or recycling container
- Because batteries can vary in size and performance, your hearing care professional is your best source for lifespan estimates and verification that you are using the proper size and type

⚠ Warnings

Batteries are dangerous if swallowed. To help prevent the accidental ingestion of batteries:

- ▲ Keep out of reach of children and pets
- Check your medications before taking them
 batteries have been mistaken for pills
- ⚠ Never put batteries in your mouth, as they can easily be swallowed
- National Button Battery Ingestion Hotline: 202-625-3333


Insertion and Removal

To insert the hearing instrument:

- 1. Hold the instrument with your thumb and forefinger on the outer edges of the case.
- Tilt your hand slightly forward and gently insert the canal tip of the instrument into your ear canal and rotate the instrument backward.
- 3. Softly press the instrument into place with your fingertip.

To remove the hearing instrument:

Grasp the instrument with your thumb and forefinger; gently rotate it as you pull outward.



3

Helpful Hints

- Minor irritation and inflammation may occur as your ear becomes accustomed to having an object in it; if so, please contact your hearing care professional.
- If an actual allergic reaction occurs, alternative materials are available; contact your hearing care professional.
- Severe swelling, discharge from the ear, excessive wax, or other unusual conditions warrant immediate consultation with a physician.

On, Off and Volume Levels With No Volume Control

☐ My hearing instrument has no volume control.
Please read the section below.
☐ My hearing instrument has a volume control.
Please skip to the section labled "On, Off and
Volume Levels with Volume Control "

To turn ON:

Insert a battery and completely close the battery door.

To turn OFF:

Open the battery door until the battery is no longer touching the battery contacts.

Your hearing system has been set to a specific volume level by your hearing care professional. If sounds are generally too loud or too soft, please contact your hearing care professional for advice and adjustment.



On, Off and Volume Levels With Volume Control

To turn the instrument "on," use your fingertip to rotate the volume control forward, toward your face.



You should feel it "click" on if the instrument was correctly turned off. To make sounds louder, continue to rotate the control forward, toward your face. Your hearing instruments may be programmed to "beep" when you reach the correct setting.

To make sounds softer, use your fingertip to rotate the control toward the back of your head.

DECREASE VOLUME OR OFF

To turn the instrument "off," rotate the control further backwards, until you feel it "click" and the control no longer turns.

Some hearing instruments are programmed to have a power on delay. Your hearing professional will place a check in the box below if your hearing instruments have this feature.

☐ My hearing instruments have a power on delay.

The power on delay means that your instruments will take several seconds to power up once they have been turned on. This feature allows you to get the hearing instruments in your ear before they start to amplify sound.

Your hearing system has been set to a specific volume level by your hearing care professional. If sounds are generally too loud or too soft, please contact your hearing care professional for advice and adjustment.

Operation

Multimemory Settings

- My hearing system has Multimemory.
 Please read the information below.
- My hearing system does not have
 Multimemory. Please go to page 16.

Your hearing care professional can set up to four hearing programs for you. These additional programs are accessed by pressing the Multimemory button inward.



When you press the Multimemory button, you may hear an indicator or a tone series indicating the device has changed to the next program. Your hearing instrument has been setup with the following environments.

Program	Tone	Speech	Environment
1	1 Tone	One	Normal
2	2 Tones	Two	
3	3 Tones	Three	
4	4 Tones	Four	

Operation

Directional Microphones

My hearing system uses omnidirectional microphones. Please skip to the Telephone Use section on page 18.
My hearing system uses directional microphones. Please read the information below.

Directional microphones help improve speech understanding in noisy situations. When directional microphones are activated automatically, the hearing instrument will continually monitor the environment and gradually adjust to the appropriate mode to optimize speech intelligibility.

Directional Settings

Your hearing system has directional microphones to help improve speech understanding in noisy situations.

Your hearing system has been setup with the following directional settings:

Program	Directional Setting		
1	Automatic	Manual	Off
2	Automatic	Manual	Off
3	Automatic	Manual	Off
4	Automatic	Manual	Off

When the directional microphones are activated automatically (automatic), the hearing instrument will continually monitor the environment and gradually adjust to the appropriate mode to optimize speech intelligibility.

When the directional microphones are activated by pushing the Multimemory button (manual), the hearing instrument optimizes speech when you select the directional memory.

Ask your hearing care professional about your particular directional settings.

Operation

Telephone Use

Your hearing instruments are equipped with tools to help you effectively communicate on the telephone. Your hearing professional will check the box/boxes that represent the telephone settings in your instruments.

Your hearing instruments have the following telephone setting(s):
☐ Automatic telephone response
☐ Automatic telecoil
☐ Telecoil. The telecoil is located in memory position
☐ A beep or indicator tone is enabled on your instruments to signal you when the device switches to the telephone setting

Automatic Telephone Response and Automatic Telecoil


These options activate the telephone response automatically when used with a hearing aid compatible telephone. To use, place the telephone receiver on your ear as you normally would and the hearing instrument will select the telephone setting. It might be necessary to move the telephone receiver slightly to find the best reception. Once the telephone is removed from the ear, the hearing instrument will switch back to the normal listening mode.

NOTE: If you have a hearing aid compatible phone, but your Automatic Telephone Response or Automatic Telecoil does not switch to the telephone setting automatically, your hearing professional can provide you with a small magnet to place on your telephone receiver. This magnet will help activate your automatic telephone option.

Operation

Telecoil and Manual Switching

Hearing instruments with a manual telecoil allow you to switch the devices to telephone mode when needed.



Manually switched telecoils are activated by choosing the telecoil setting using the Multimemory button.



Some hearing instruments work best by holding the phone close to, but not fully covering your ear. In some instances, if you encounter whistling (feedback), tilt the receiver at an angle until the whistling stops. Your hearing professional can give instructions and techniques for your specific needs.





Some hearing aid users have reported a buzzing sound in their hearing aids when they are using cell phones, indicating that the cell phone and hearing aid may not be compatible. According to the ANSI C63.19 standard (ANSI C63.19-2006 American National Standard Methods of Measurement of Compatibility Between Wireless Communications Devices and Hearing Aids), the compatibility of a particular hearing aid and cell phone can be predicted by adding the rating for the hearing aid immunity to the rating for the cell phone emissions. For example, the sum of a hearing aid rating of 2 (M2/T2) and a telephone rating of 3 (M3/T3) would result in a combined rating of 5. Any combined rating that equals at least 5 would provide "normal use"; a combined rating of 6 or greater would indicate "excellent performance."

The immunity of this hearing aid is at least M2/T2. The equipment performance measurements, categories and system classifications are based upon the best information available but cannot guarantee that all users will be satisfied.

NOTE: The performance of individual hearing aids may vary with individual cell phones. Therefore, please try this hearing aid with your cell phone

Operation

or, if you are purchasing a new phone, be sure to try it with your hearing aid prior to purchase. For additional guidance, please ask your cell phone provider for the booklet entitled "Hearing Aid Compatibility with Digital Wireless Cell Phones."

Self Check

Some hearing systems can run a diagnostic test of their own performance. This feature is called Self Check. Your hearing care professional will check the box below if this feature is activated.

$\ \square$ Your Self Check feature is activated.
Your Self Check indicator is a:
☐ Speech Indicator (Good, Make Appointment)
☐ Tone Indicator

To perform the Self Check:

- 1. Open and close your battery door three times.
- 2. Place the hearing system in your ear.
- 3. Listen for the status indicator.
- 4. Schedule an appointment with your hearing care professional, if needed.

Reminder

Some hearing systems have been programmed with a Reminder indicating that it is time to schedule an appointment. Your hearing care professional will check the box below if this feature is activated. The Reminder will occur automatically. When you hear the Reminder indicator, please contact your hearing care professional.

☐ Your Reminder feature is activated.
Your Reminder indicator is a:
☐ Speech Indicator (Make Appointment)
☐ Tone Indicator

Operation

Summary of Settings

	Environment	Directional Setting
Program 1	Normal	□ Off □ Automatic
Program 2		□ Off □ Automatic
Program 3		□ Off □ Automatic □ Manual
Program 4		□ Off □ Automatic □ Manual
Low Battery	_	_
Battery End of Life	_	_
Standby	_	_
Push and Hold	_	_
Self Check		_
Reminder	_	_

Indicator Tone On	Indicator Tone Type
□ Yes □ No	□ Speech □ 1 Tone
□ Yes □ No	□ Speech □ 2 Tones □ Tone Series
□ Yes □ No	☐ Speech ☐ 3 Tones ☐ Tone Series
□ Yes □ No	☐ Speech ☐ 4 Tones ☐ Tone Series
□ Yes □ No	□ Speech □ 4 Tones □ Tone Series
□ Yes □ No	☐ Speech ☐ Sustained Tone
□ Yes □ No	☐ Tone Series
□ Yes □ No	☐ Speech ☐ Tone Series
□ Yes □ No	☐ Speech ☐ Tone Series
□ Yes □ No	□ Speech □ Tone Series

Hearing System Care

Instrument Care

Do your best to keep your hearing instrument clean at all times. Heat, moisture and foreign substances can result in poor performance.

- Clean daily over a soft cloth to prevent damage from a fall to a hard surface
- Use a cleaning brush to clean debris from around the microphone, receiver and the battery compartment
- Never use water, solvents, cleaning fluids or oil to clean your instrument

Your hearing care professional can provide further information on additional maintenance procedures for your hearing system if needed.

Helpful Hints

- When not wearing your hearing instruments, open the battery door to allow any moisture to evaporate
- When not in use, remove the batteries completely; place your hearing system in the storage container and store:
 - In a dry, safe place
 - Away from direct sunlight or heat to avoid extreme temperatures
 - Where you can easily find them
 - Safely out of reach of pets and children
- Do not take apart your hearing instruments or insert the cleaning tools inside them

Hearing System Care

Service and Repair

If, for any reason, your hearing system does not operate properly, do NOT attempt to fix it yourself. Not only are you likely to violate any applicable warranties or insurance, you could easily cause further damage.

Should your hearing system fail or perform poorly, check the guide on the next page for possible solutions. If problems continue, contact your hearing care professional for advice and assistance. He or she is able to solve many common problems right in his or her office or clinic.

Troubleshooting Guide

SYMPTOM	POSSIBLE CAUSES	SOLUTIONS
	Low battery	Replace battery
Not Loud Enough	Wax or debris in the microphone or receiver	Clean both microphone and receiver with brush
	Hearing change	Contact your hearing care professional
Inconsistent Performance	Low battery	Replace battery
Unclear,	Low battery	Replace battery
Distorted Performance	Defective hearing instrument	Contact your hearing care professional
	Low battery	Replace battery
Dead	Wax or debris in the microphone or receiver	Clean both microphone and receiver with brush

Tips for Better Communication

Your hearing care professional will recommend an appropriate schedule to help you adapt to your new hearing system. It will take practice, time and patience for your brain to adapt to the new sounds that your hearing system provides. Hearing is only part of how we share thoughts, ideas and feelings. Reading lips, facial expressions and gestures can help the learning process and add to what amplification alone may miss.

Please review the following simple communication tips:

For You

- Move closer to and look at the speaker
- Sit face-to-face in a quiet room
- Try different locations to find the best place to listen
- Minimize distractions
- Background noises may be frustrating at first; remember, you have not heard them for a while
- Let others know what you need; keep in mind that people cannot "see" your hearing loss

Tips for Better Communication

- Develop realistic expectations of what your hearing instruments can and cannot do
- Better hearing with hearing instruments is a learned skill combining desire, practice and patience

For Your Family and Friends

Your family and friends are also affected by your hearing loss. Request that they:

- Get your full attention before beginning to speak
- Look at you or sit face-to-face in a quiet room
- Speak clearly and at a normal rate and level; shouting can actually make understanding more difficult
- Rephrase rather than repeat the same words; different words may be easier to understand
- Minimize distractions while speaking

FDA Information

Required Information

The following additional information is provided in compliance with U.S. Food and Drug Administration (FDA) regulations:

WARNING TO HEARING AID DISPENSERS. A hearing aid dispenser should advise a prospective hearing aid user to consult promptly with a licensed physician (preferably an ear specialist) before dispensing a hearing aid if the hearing aid dispenser determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- i. Visible congenital or traumatic deformity of the ear.
- ii. History of active drainage from the ear within the previous 90 days.
- History of sudden or rapidly progressive hearing loss within the previous 90 days.
- iv. Acute or chronic dizziness.
- v. Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- vi. Audiometric air-bone gap equal to or greater than 15 decibels at 500 Hertz (Hz), 1,000 Hz and 2,000 Hz.
- vii. Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- viii. Pain or discomfort in the ear.

RF IMMUNITY LEVEL. This hearing instrument has a cell phone immunity rating of M2/T2. A hearing aid compatible cell phone must carry a rating of M3/T3 or higher to work with this hearing aid. Please consult your cell phone specification for the cell phone immunity rating.

IMPORTANT NOTICE FOR PROSPECTIVE HEARING AID USERS.

Good health practice requires that a person with a hearing loss have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before purchasing a hearing aid. Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists, or otorhynolaryngologists. The purpose of the medical evaluation is to assure that all medically treatable conditions that may affect hearing are identified and treated before the hearing aid is purchased.

Following the medical evaluation, the physician will give you a written statement that states that your hearing loss has been medically evaluated and that you may be considered a candidate for a hearing aid. The physician will refer you to an audiologist or hearing aid dispenser, as appropriate, for a hearing aid evaluation.

The audiologist or hearing aid dispenser will conduct a hearing aid evaluation to assess your ability to hear with and without a hearing aid. The hearing aid evaluation will enable the audiologist or dispenser to select and fit a hearing aid to your individual needs.

If you have reservations about your ability to adapt to amplification, you should inquire about the availability of a trial-rental or purchase-option program. Many hearing aid dispensers now offer programs that permit you to wear a hearing aid for a period of time for a nominal fee after which you may decide if you want to purchase the hearing aid.

Federal law restricts the sale of hearing aids to those individuals who have obtained a medical evaluation from a licensed physician. Federal law permits a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best health interest and its use is strongly discouraged.

A hearing aid will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions. In most cases infrequent use of a hearing aid does not permit a user to attain full benefit from it.

CHILDREN WITH HEARING LOSS. In addition to seeing a physician for a medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation since hearing loss may cause problems in language development and the educational and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with a hearing loss.

NuEar 6769 Mesa Ridge Road Suite 100 San Diego, CA 92121 800.626.8327 858.450.9972

www.NuEar.com



